



CERTIFICATE OF ANALYSIS

Potassium Eluant For Biogenic Amines Analysis

K563

Product: Potassium Eluant for Biogenic Amines Analysis

Part Number: K563

Lot Number: 201005

Production Date: 04 Jan 2012

Expiration Date: 04 Jan 2014

Primary Application: Eluant for ALKION Cation-exchange chromatography of polyamines.

Acceptance Criteria:

1. pH = $5.63 \pm 0.10 @ 25 \pm 0.4 ^\circ\text{C}$ (Y/N)

2. Peak resolution meets specification (Y/N)

3. Baseline: no peaks (Y/N)

Y
Y
Y

Storage Conditions: Dry at 8 to 25°C

Analysis Certified by:

John Murphy
VP OPER
Title

Date:

19 MAR 2012

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 6, 04/18/2011