CERTIFICATE OF ANALYSIS

Potassium Eluant For
Biogenic Amines Analysis

K563

Product: Potassium Eluant for Biogenic Amines Analysis
Part Number: K563
Lot Number: 2103009
Production Date: March 3, 2021
Expiration Date: March 3, 2023

Primary Application: Eluant for ALKION Cation-exchange chromatography of polyamines.

Acceptance Criteria:
1. pH = 5.63 ± 0.10 @ 25 ± 0.4 ˚C (Y/N) **Y**

Storage Conditions: Dry at 8 to 25ºC

Analysis Certified by: Jim Murphy **Date:** April 20, 2021
President
Title

Rev. B, 9/16/15

1280 Space Park Way · Mountain View CA 94043 · USA
800-654-3330 · 650-694-6700 · 650-968-0749 FAX · www.pickeringlabs.com · support@pickeringlabs.com