



## CERTIFICATE OF ANALYSIS

### Potassium Eluant For Biogenic Amines Analysis

**K600**

**Product:** Potassium Eluant for Biogenic Amines Analysis

**Part Number:** K600

**Lot Number:** 111026

**Production Date:** 17 Nov 2011

**Expiration Date:** 16 Nov 2013

**Primary Application:** Eluant for ALKION Cation-exchange chromatography of polyamines.

**Acceptance Criteria:**

1. pH = 6.00 ± 0.10 @ 25 ± 0.4 °C (Y/N)
2. Chromatogram with well resolved peaks (Y/N)

4  
4

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:**

*Jim Murphy*  
*VP Ops*  
Title

**Date:** 21 Nov 2011

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 5, 8/15/08