



CERTIFICATE OF ANALYSIS

Potassium Eluant For Biogenic Amines Analysis

K600

Product: Potassium Eluant for Biogenic Amines Analysis

Part Number: K600

Lot Number: 2601018

Production Date: 12 Jan 2026

Expiration Date: 12 Jan 2028

Primary Application: Eluant for ALKION Cation-exchange chromatography of polyamines.

Acceptance Criteria:

1. pH = 6.00 ± 0.10 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: Dry at 8 to 25°C

Analysis Certified by: David Mazawa
Title: QC Manager

Date: 19 Jan 2026