



CERTIFICATE OF ANALYSIS
Lithium Diluent for Amino Acid Analysis

Li220

Product: Lithium Diluent for Amino Acid Analysis

Part Number: Li220

Lot Number: 004008

Test Start Date: 09 Apr 2010

Expiration Date: 09 Apr 2011

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.36 ± 0.01 @ 25 ± 0.4 °C
2. Baseline: no peaks
3. Conductivity after adding lithium citrate 6.08 ± 0.01 mS

2.365
4
4

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Murphy
VP Ops

Title

Date: _____

21 Apr 2010

Rev. 5, 8/11/08

ACCEPTED
PICKERING LABORATORIES, INC
QUALITY ASSURANCE