



CERTIFICATE OF ANALYSIS

Lithium Diluent for Amino Acid Analysis

Li220

Product: Lithium Diluent for Amino Acid Analysis

Part Number: Li220

Lot Number: 2401026

Production Date: 22 Jan 2024

Expiration Date: 22 Jan 2026

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.36 ± 0.01 @ 25 ± 0.4 °C (Y/N)

Y

2. Baseline: no peaks (Y/N)

Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: David Mazawa

Date: 24 Jan 2024

Title: QC Manager