



## CERTIFICATE OF ANALYSIS

### Lithium Diluent for Amino Acid Analysis

**Li220**

**Product:** Lithium Diluent for Amino Acid Analysis

**Part Number:** Li220

**Lot Number:** 2404021

**Production Date:** 11 April 2024

**Expiration Date:** 11 April 2026

**Primary Application:** Lithium-based amino acid analysis.

**Acceptance Criteria:**

1. pH =  $2.36 \pm 0.01$  @  $25 \pm 0.4$  °C (Y/N)

Y

2. Baseline: no peaks (Y/N)

Y

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** David Mazawa

**Date:** 23 April 2024

**Title:** QC Manager