



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis

Li275

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li275

Lot Number: 002013

Test Start Date: 10 Feb 2010

Expiration Date: 10 Feb 2011

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.75 ± 0.01 @ 25 ± 0.4 °C
2. Resolution:
 - Threonine/Serine > 72% (Y/N)
3. Elution time of L-Aspartic acid meets specifications (Y/N)
4. Conductivity: 6.08 ± 0.01 mS

Actual Values

2.750

Yes.

Yes.

6.08

Calibration Standard:

Lithium Calibration Standard, PN 011006p, LN 712009

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: [Signature]

Date: 02 Mar 2010

VP OPR
Title

Rev. 6, 8/11/08

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE