



CERTIFICATE OF ANALYSIS

Lithium Eluant (pH 2.75) for Amino Acid Analysis

Li275LC

Product: Lithium Eluant pH2.75 for Amino Acid Analysis

Part Number: Li275LC

Lot Number: 08012

Test Start Date: 22 AUG 2011

Expiration Date: 18 AUG 2013

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.75 ± 0.01 @ 25 ± 0.4 °C
2. Resolution:
 - Threonine/Serine > 72% (Y/N)
3. Elution time of L-Aspartic acid meets specifications (Y/N)

Actual Values

Y

Y

Y

Calibration Standard:

Lithium Calibration Standard, PN 011006p, LN 909011

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Murphy
VP OPR
Title

Date: 20 Sep 2011

ACCEPTED

PICKERING LABORATORIES, INC
QUALITY ASSURANCE

Rev. 6, 8/11/08