



CERTIFICATE OF ANALYSIS

Lithium Eluant (pH3.75) for Amino Acids Analysis

Li375

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li375

Lot Number: 202040

Production Date: 01 MAR 2012

Expiration Date: 28 FEB 2014


Primary Application: Lithium-based amino acids analysis.

Acceptance Criteria:

- 1. pH = 3.75 ± 0.01 @ 25 ± 0.4 °C (Y/N) y
- 2. Resolution meets specifications: y
 - Tryptophan, Ethanolamine and Ammonia (Y/N) y
 - Lysine, 1-Methylhistidine and Histidine (Y/N) y
- 3. Elution time of Histidine and Arginine meets specification (Y/N) y
- 4. Baseline: no peaks (Y/N) y

Calibration Standard: Lithium Calibration Standard, PN: 1700-0170

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  **Date:** 05 MAR 2012
Title

ACCEPTED

Rev. 11, 10/07/2011

PICKERING LABORATORIES, INC
QUALITY ASSURANCE

P:\Documentation\Chemical Documentation\Li375