



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis Li750

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li750

Lot Number: 103022

Production Date: 14 MAR 2011

Expiration Date: 11 MAR 2012

Primary Application: Lithium amino acid analysis


Acceptance Criteria:

- 1. pH = 7.50 ± 0.15 @ 25 ± 0.4 °C (Y/N) y
- 2. Resolution:
 - All peaks in the Lithium Calibration Standard must be clearly resolved. (Y/N) y
- 3. Elution time: (For Agilent 1100 HPLC System)
 - Histidine = 97.4 ± 0.5 min. y
- 4. Baseline: no peaks (Y/N) y

Calibration Standard:

Lithium Calibration Standard, PN 011006P,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  **Date:** 21 MAR 2011

VP Operator
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE