



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis Li750

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li750

Lot Number: 104006

Production Date: 08 Apr 2011

Expiration Date: 07 Apr 2012

Primary Application: Lithium amino acid analysis

Acceptance Criteria:

1. pH = 7.50 ± 0.15 @ 25 ± 0.4 °C (Y/N) Y
2. Resolution: Y
 - All peaks in the Lithium Calibration Standard must be clearly resolved. (Y/N)
3. Elution time: (For Agilent 1100 HPLC System) Y
 - Histidine = 97.4 ± 0.5 min.
4. Baseline: no peaks (Y/N) Y

Calibration Standard:

Lithium Calibration Standard, PN 011006P,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: *Jim Murphy* Date: 13 Apr 2010
VP OPR
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE