



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis Li750

Product: Lithium Eluant for Amino Acid Analysis
Part Number: Li750
Lot Number: 105002
Production Date: 05 MAY 2011
Expiration Date: 03 MAY 2012


Primary Application: Lithium amino acid analysis

Acceptance Criteria:

1. pH = 7.50 ± 0.15 @ 25 ± 0.4 °C (Y/N) Y
2. Resolution:
 - All peaks in the Lithium Calibration Standard must be clearly resolved. (Y/N) Y
3. Elution time: (For Agilent 1100 HPLC System)
 - Histidine = 94.0 ± 0.5 min. (Y/N) Y
4. Baseline: no peaks (Y/N) Y

Calibration Standard:
Lithium Calibration Standard, PN 011006P,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  Date: 10 MAY 2011
Title: VP OPER

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 10, 04/08/2011