



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis Li750

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li750

Lot Number: 109038

Production Date: 29 Sep 2011

Expiration Date: 28 Sep 2012

Primary Application: Lithium amino acid analysis

Acceptance Criteria:

1. pH = 7.50 ± 0.15 @ 25 ± 0.4 °C (Y/N) y

2. Resolution:

- All peaks in the Lithium Calibration Standard must be clearly resolved. (Y/N) y

3. Elution time: Histidine = 94.0 ± 1.0 min. (Y/N) y

4. Baseline: no peaks (Y/N) y

Calibration Standard:

Lithium Calibration Standard, PN: 011006P,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: *[Signature]*

Date: 05 Oct 2011

VP Ops
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 10, 04/08/2011