



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis Li750

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li750

Lot Number: 202017

Production Date: 16 Feb 2012

Expiration Date: 14 Feb 2013

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 7.50 ± 0.15 @ 25 ± 0.4 °C (Y/N)

2. Resolution:

- All peaks in the Lithium Calibration standard are clearly resolved (Y/N)

3. Elution time of Histidine meets specifications (Y/N)

4. Baseline: no peaks (Y/N)

Y
Y
Y
Y

Sample: Lithium Calibration Standard, PN: 011006P

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

Jim Murphy
VP OPER
Title

Date:

29 FEB 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 10, 09/23/2011

P:\Documentation\Chemical Documentation\Li750