



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis Li750

Product: Lithium Eluant for Amino Acid Analysis

Part Number: Li750

Lot Number: 203001

Production Date: 01 MAR 2012

Expiration Date: 01 MAR 2013

Primary Application: Lithium-based amino acid analysis.

Acceptance Criteria:

1. pH = 7.50 ± 0.15 @ 25 ± 0.4 °C (Y/N) Y

2. Resolution: Y

- All peaks in the Lithium Calibration standard are clearly resolved (Y/N) Y

3. Elution time of Histidine meets specifications (Y/N) Y

4. Baseline: no peaks (Y/N) Y

Sample: Lithium Calibration Standard, PN: 011006P

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: [Signature]

VP Oper
Title

Date: 06 MAR 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 10, 09/23/2011

P:\Documentation\Chemical Documentation\Li750