PICKERING
LABORATORIES

CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis
Li750

Product: Lithium Eluant for Amino Acid Analysis
Part Number: Li750
Lot Number: 107010
Production Date: 25 Jul 2011
Expiration Date: 20 Jul 2012

Primary Application: Lithium amino acid analysis

Acceptance Criteria:

1. pH = 7.50 ± 0.15 @ 25 ± 0.4 °C(Y/N) — Y
2. Resolution:
   - All peaks in the Lithium Calibration Standard must be clearly resolved. (Y/N) — Y
3. Elution time: (For Agilent 1100 HPLC System)
   - Histidine = 94.0 ± 0.5 min.(Y/N) — Y
4. Baseline: no peaks (Y/N) — Y

Calibration Standard:
Lithium Calibration Standard, **PN** 011006P,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: [signature] **Date:** 29 Jul 2011

VP Opr
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 10, 04/08/2011

1280 Space Park Way · Mountain View CA 94043 · USA
800-654-3330 · 650-694-6700 · 650-968-0749 FAX · www.pickeringlabs.com · support@pickeringlabs.com