CERTIFICATE OF ANALYSIS

Sodium Diluent for Amino Acid Analysis

Na220

Product: Sodium Diluent for Amino Acid Analysis
Part Number: Na220
Lot Number: 004025
Production Date: April 16, 2020
Expiration Date: April 16, 2022

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.20 ± 0.01 @ 25 ± 0.4 ˚C (Y/N)	**Y**
2. Baseline: no peaks (Y/N)	**Y**

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Rebecca Smith **Date:** April 29, 2020
Vice President
Title

1280 Space Park Way · Mountain View CA 94043 · USA
800-654-3330 · 650-694-6700 · 650-968-0749 FAX · www.pickeringlabs.com · support@pickeringlabs.com