CERTIFICATE OF ANALYSIS

Sodium Diluent for Amino Acid Analysis

Na220

Product: Sodium Diluent for Amino Acid Analysis
Part Number: Na220
Lot Number: 006019
Production Date: June 9, 2020
Expiration Date: June 9, 2022

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.20 ± 0.01 @ 25 ± 0.4 ˚C (Y/N) **Y**

2. Baseline: no peaks (Y/N) **Y**

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Rebecca Smith **Date:** July 1, 2020
Vice President
Title

Rev. A, 10/24/12 **P:\Documentation\Chemical Documentation\Na220**

1280 Space Park Way · Mountain View CA 94043 · USA
800-654-3330 · 650-694-6700 · 650-968-0749 FAX · www.pickeringlabs.com · support@pickeringlabs.com