



CERTIFICATE OF ANALYSIS

Sodium Diluent for Amino Acid Analysis

Na220

Product: Sodium Diluent for Amino Acid Analysis

Part Number: Na220

Lot Number: 010019

Test Start Date: 20 Oct 2010

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.20 ± 0.01 @ 25 ± 0.4 °C

2. No peaks in baseline.

Actual Values

2.206

4

Analysis Certified by: *Jim Murphy* **Date:** 27 Oct 2010

VP OPR
Title

ACCEPTED

PICKERING LABORATORIES, INC
QUALITY ASSURANCE