



CERTIFICATE OF ANALYSIS

Sodium Diluent for Amino Acid Analysis

Na220

Product: Sodium Diluent for Amino Acid Analysis

Part Number: Na220

Lot Number: 101027

Test Start Date: 24 JAN 2011

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.20 ± 0.01 @ 25 ± 0.4 °C

2. No peaks in baseline.

Actual Values

2.199

4

Analysis Certified by: *Jim Murphy* Date: 15 FEB 2011

VP Ops

Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE