



## CERTIFICATE OF ANALYSIS

### Sodium Diluent for Amino Acid Analysis

**Na220**

**Product:** Sodium Diluent for Amino Acid Analysis

**Part Number:** Na220

**Lot Number:** 2206032

**Production Date:** 24 June 2022

**Expiration Date:** 24 June 2024

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

1. pH =  $2.20 \pm 0.01$  @  $25 \pm 0.4$  °C (Y/N) Y

2. Baseline: no peaks (Y/N) Y

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** David Mazawa **Date:** 24 June 2022  
**Title** Technical Support Chemist