



CERTIFICATE OF ANALYSIS

Sodium Diluent for Amino Acid Analysis

Na220L

Product: Sodium Diluent for Amino Acid Analysis

Part Number: Na220L

Lot Number: 2509014

Production Date: 04 Sep 2025

Expiration Date: 04 Sep 2027

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.20 ± 0.01 @ 25 ± 0.4 °C (Y/N)

Y

2. Baseline: no peaks (Y/N)

Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: David Mazawa
Title QC Manager

Date: 10 Sep 2025