



# CERTIFICATE OF ANALYSIS

## Sodium Eluant for Amino Acid Analysis

Na270

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na270

Lot Number: 008014

Test Start Date: 11 AUG 2010

Expiration Date: 12 AUG 2011

Primary Application: Sodium-based amino acid analysis.

### Acceptance Criteria:

1. pH =  $2.80 \pm 0.01$  @  $25 \pm 0.4$  °C
2. Sample 2<sup>nd</sup>, 3<sup>rd</sup>, 4<sup>th</sup>, and 5<sup>th</sup> peaks touch the bottom of baseline (Y/N)
3. No peaks on baseline (Y/N)

### Actual Values

2.794

Y

Y

Calibration Standard: Oxidized Hydrolysate, PN 1700-0155, LN 309018

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  Date: 23 Aug 2010

Title

Rev.7, 8/11/08

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE