

CERTIFICATE OF ANALYSIS

Sodium Eluant for Amino Acid Analysis

Na270

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na270

Lot Number: 102009

Test Start Date: 11 Feb 2011

Expiration Date: 10 Feb 2013

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.80 ± 0.01 @ 25 ± 0.4 °C
2. Sample 2nd, 3rd, 4th, and 5th peaks touch the bottom of baseline (Y/N)
3. No peaks on baseline (Y/N)

Actual Values

2.807

Y

Y

Calibration Standard: Oxidized Hydrolysate, PN 1700-0155, LN 309018

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Murphy
VP Oper

Title

Date: _____

17 Feb 2011

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE