

CERTIFICATE OF ANALYSIS
Sodium Eluant for Amino Acid Analysis

Na270

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na270

Lot Number: 103036

Production Date: 31 MAR 2011

Expiration Date: 31 MAR 2013

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 2.80 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Sample 2nd, 3rd, 4th, and 5th peaks touch the bottom of baseline (Y/N)
3. No peaks on baseline (Y/N)

Y
Y
Y

Calibration Standard: Oxidized Hydrolysate, PN 1700-0155

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: [Signature]

Date: 12 Apr 2011

V.P. Orr
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE