



# CERTIFICATE OF ANALYSIS

## Sodium Eluant for Amino Acid Analysis

Na270

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na270

Lot Number: 106 014

Production Date: 15 JUN 2011

Expiration Date: 14 JUN 2013

Primary Application: Sodium-based amino acid analysis.

### Acceptance Criteria:

1. pH =  $2.80 \pm 0.01$  @  $25 \pm 0.4$  °C (Y/N)
2. Well resolved 2<sup>nd</sup>, 3<sup>rd</sup>, 4<sup>th</sup>, and 5<sup>th</sup> peaks (Y/N)
3. No peaks on baseline (Y/N)

Y  
Y  
Y

Calibration Standard: Oxidized Hydrolysate, PN 1700-0155

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: *Jim Murphy* Date: 21 JUN 2011

VP Opk  
Title

**ACCEPTED**

PICKERING LABORATORIES, INC  
QUALITY ASSURANCE

Rev.8, 04/12/2011