



CERTIFICATE OF ANALYSIS

Sodium Eluant (pH 3.15) For Amino Acids Analysis

Na315

Product: Sodium Eluant pH 3.15 for Amino Acid Analysis

Part Number: Na315

Lot Number: 109031

Production Date: 20 Sep 2011

Expiration Date: 19 Sep 2013

Primary Application: Sodium eluant for Sodium-based amino acid analysis.

Acceptance Criteria:

1. Retention time of L-aspartic acid meets specifications (Y/N)
2. Resolution of Thr/Ser \geq 93% (Y/N)
3. Height of ammonia plateau/height of ammonia peak \leq 9% (Y/N)
4. No peaks in baseline (Y/N)

Y
Y
Y
Y

Sample: Collagen hydrolysate, PN 012506C,

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: [Signature]

Date: 21 Sep 2011

Title: V.P. OPR

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 11, 2/24/2011