



# CERTIFICATE OF ANALYSIS

## Sodium Eluant (pH 3.15) For Amino Acids Analysis

### Na315

**Product:** Sodium Eluant pH 3.15 for Amino Acid Analysis

**Part Number:** Na315

**Lot Number:** 111010

**Production Date:** 07 Nov 2011

**Expiration Date:** 07 Nov 2013

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

- 1. pH = 3.15 ± 0.01 @ 25 ± 0.4 °C (Y/N)
- 2. Resolution meets specifications:
  - Threonine/Serine (Y/N)
- 3. Elution time of L-Aspartic acid meets specifications (Y/N)
- 4. Baseline no peaks (Y/N)

Y  
Y  
Y  
Y

**Sample:** Collagen hydrolysate, PN: 012506C

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:**

*[Signature]*  
 VP OPR  
 TIME

**Date:**

21 Nov 2011

Rev. 12, 11/09/2011

# ACCEPTED

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

P:\Documentation\Chemical Documentation\Na315