



CERTIFICATE OF ANALYSIS

Sodium Eluant (pH 3.15) For Amino Acids Analysis

Na315

Product: Sodium Eluant pH 3.15 for Amino Acid Analysis

Part Number: Na315

Lot Number: 203034

Production Date: 28 MAR 2012

Expiration Date: 27 MAR 2014

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 3.15 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Resolution meets specifications:
 - Threonine/Serine (Y/N)
3. Elution time of L-Aspartic acid meets specifications (Y/N)
4. Baseline: no peaks (Y/N)

Y
Y
Y
Y

Sample: Collagen hydrolysate, PN: 012506C

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

Jim Murphy
VP OPR
Title

Date:

06 APR 2012

ACCEPTED

Rev. 12, 11/09/2011

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

P:\Documentation\Chemical Documentation\Na315