



## CERTIFICATE OF ANALYSIS

### Sodium Eluant (pH 3.15) For Amino Acids Analysis

**Na315**

**Product:** Sodium Eluant pH 3.15 for Amino Acid Analysis

**Part Number:** Na315

**Lot Number:** 2304032

**Production Date:** 20 April 2023

**Expiration Date:** 20 April 2025

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

- |  |          |
|--|----------|
| 1. pH = $3.15 \pm 0.01$ @ $25 \pm 0.4$ °C (Y/N)  | <u>Y</u> |
| 2. Resolution meets specifications: <ul style="list-style-type: none"><li>• Threonine/Serine (Y/N)</li></ul> | <u>Y</u> |
| 3. Elution time of L-Aspartic acid meets specifications (Y/N)  | <u>Y</u> |
| 4. Baseline: no peaks (Y/N)  | <u>Y</u> |

**Sample:** Collagen hydrolysate, **PN:** 012506C

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** David Mazawa **Date:** 24 April 2023  
**Title:** Technical Support Chemist