



CERTIFICATE OF ANALYSIS

Sodium Eluant (pH 3.15) For Amino Acids Analysis

Na315

Product: Sodium Eluant pH 3.15 for Amino Acid Analysis

Part Number: Na315

Lot Number: 304022

Production Date: 24 April 2013

Expiration Date: 24 April 2015

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 3.15 ± 0.01 @ 25 ± 0.4 °C (Y/N) y

2. Resolution meets specifications:
• Threonine/Serine (Y/N) y

3. Elution time of L-Aspartic acid meets specifications (Y/N) y

4. Baseline: no peaks (Y/N) y

Sample: Collagen hydrolysate, **PN:** 012506C

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** 28 May 2013
President
Title