



## CERTIFICATE OF ANALYSIS

### Sodium Eluant (pH 3.15) For Amino Acids Analysis

**Na315**

**Product:** Sodium Eluant pH 3.15 for Amino Acid Analysis

**Part Number:** Na315

**Lot Number:** 503002

**Production Date:** 02 Mar 2015

**Expiration Date:** 02 Mar 2017

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

- |   |          |
|---|----------|
| 1. pH = 3.15 ± 0.01 @ 25 ± 0.4 °C (Y/N)                       | <u>y</u> |
| 2. Resolution meets specifications:                           |          |
| • Threonine/Serine (Y/N)                                      | <u>y</u> |
| 3. Elution time of L-Aspartic acid meets specifications (Y/N) | <u>y</u> |
| 4. Baseline: no peaks (Y/N)                                   | <u>y</u> |

**Sample:** Collagen hydrolysate, **PN:** 012506C

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** 05 Mar 2015  
President  
Title