



CERTIFICATE OF ANALYSIS

Sodium Eluant (pH 3.15) For Amino Acids Analysis

Na315

Product: Sodium Eluant pH 3.15 for Amino Acid Analysis

Part Number: Na315

Lot Number: 704031

Production Date: 24 Apr 2017

Expiration Date: 24 Apr 2019

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

- | | |
|---|----------|
| 1. pH = 3.15 ± 0.01 @ 25 ± 0.4 °C (Y/N) | <u>Y</u> |
| 2. Resolution meets specifications: | |
| • Threonine/Serine (Y/N) | <u>Y</u> |
| 3. Elution time of L-Aspartic acid meets specifications (Y/N) | <u>Y</u> |
| 4. Baseline: no peaks (Y/N) | <u>Y</u> |

Sample: Collagen hydrolysate, **PN:** 012506C

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Tony McIsaac **Date:** 28 Apr 2017
VP Manufacturing
Title