

CERTIFICATE OF ANALYSIS

Sodium Eluant for Amino Acid Analysis

Na328

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na328

Lot Number: 104023

Production Date: 18 APR 2011

Expiration Date: 14 APR 2013

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 3.28 ± 0.01 @ 25 ± 0.4 °C (Y/N) 4

2. Retention time of L-aspartic acid = 8.5 ± 0.5 minutes (Y/N) 4

3. No peaks on baseline (Y/N) 4

Calibration Standard: Protein Hydrolysate, PN 012506H

Analysis Certified by: *[Signature]*

Date: 25 Apr 2011

VP Ops
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE