



## CERTIFICATE OF ANALYSIS

**Sodium Eluant (pH 3.28) For Amino Acids Analysis**

**Na328**

**Product:** Sodium Eluant pH 3.28 for Amino Acid Analysis

**Part Number:** Na328

**Lot Number:**

112006

**Production Date:**

08 Dec 2011

**Expiration Date:**

07 Dec 2013

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

1. pH =  $3.28 \pm 0.01$  @  $25 \pm 0.4$  °C (Y/N)
2. Elution time of L-Aspartic acid meets specifications (Y/N)
3. Baseline: no peaks (Y/N)

Y  
Y  
Y

**Sample:** Protein Hydrolysate, **PN:** 012506H

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:**

*[Signature]*  
VP OPR  
Title

**Date:**

05 MAR 2012

**ACCEPTED**

Rev. 07, 11/09/2011

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

P:\Documentation\Chemical Documentation\Na328