



## CERTIFICATE OF ANALYSIS

### Sodium Eluant (pH 3.28) For Amino Acids Analysis

**Na328**

**Product:** Sodium Eluant pH 3.28 for Amino Acid Analysis

**Part Number:** Na328

**Lot Number:** 906002

**Production Date:** June 3, 2019

**Expiration Date:** June 3, 2021

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

- |   |          |
|---|----------|
| 1. pH = $3.28 \pm 0.01$ @ $25 \pm 0.4$ °C (Y/N)               | <u>Y</u> |
| 2. Elution time of L-Aspartic acid meets specifications (Y/N) | <u>Y</u> |
| 3. Resolution of Threonine-Serine meets specification (Y/)    | <u>Y</u> |
| 4. Baseline: no peaks (Y/N)                                   | <u>Y</u> |

**Sample:** Protein Hydrolysate, **PN:** 012506H

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Rebecca Smith **Date:** 6/7/2019  
Director of Operations  
Title