



**CERTIFICATE OF ANALYSIS**  
**Sodium Eluant for Amino Acid Analysis**  
**Na425**

**Product:** Sodium Eluant for Amino Acid Analysis

**Part Number:** Na425

**Lot Number:** 009013

**Test Start Date:** 07 Sep 2010

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

1. pH =  $4.25 \pm 0.01$  @  $25 \pm 0.4$  °C

**Actual Values**

4.254

2. Peak resolution meets specifications (Y/N).

Y

**Calibration Standard:** Protein Hydrolysate, PN 012506H, LN 709037

**Analysis Certified by:** \_\_\_\_\_

*John Murphy*  
VT Opr

Title

**Date:** 17 Sep 2010

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE