



**CERTIFICATE OF ANALYSIS**  
**Sodium Eluant pH 4.25 for Amino Acid Analysis**

**Na425**

**Product:** Sodium Eluant for Amino Acid Analysis

**Part Number:** Na425

**Lot Number:** 009043

**Production Date:** September 21, 2020

**Expiration Date:** September 21, 2022

**Acceptance Criteria:**

1. pH = 4.25 ± 0.01 @ 25 ± 0.4 °C (Y/N)

Y

2. Resolution meets specifications:

• Alanine/Cystine (Y/N)

Y

• Isoleucine/Leucine (Y/N)

Y

3. Elution time of Valine and Methionine meets specification (Y/N)

Y

4. Baseline: no peaks (Y/N)

Y

**Calibration Standard:** Protein Hydrolysate Standard, **PN:** 012506H

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy

**Date:** October 15, 2020

President

**Title**

Rev. A, 10/24/12

P:\Documentation\Chemical Documentation\Na425