



CERTIFICATE OF ANALYSIS

Sodium Eluant pH 4.25 for Amino Acid Analysis

Na425

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na425

Lot Number: 2506047

Production Date: 25 Jun 2025

Expiration Date: 25 Jun 2027

Acceptance Criteria:

1. pH = 4.25 ± 0.01 @ 25 ± 0.4 °C (Y/N)

Y

2. Resolution meets specifications:

• Alanine/Cystine (Y/N)

Y

• Isoleucine/Leucine (Y/N)

Y

3. Elution time of Valine and Methionine meets specification (Y/N)

Y

4. Baseline: no peaks (Y/N)

Y

Calibration Standard: Protein Hydrolysate Standard, **PN:** 012506H

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: David Mazawa

Date: 03 Jul 2025

Title: QC Manager