



CERTIFICATE OF ANALYSIS
Sodium Eluant pH 4.25 for Amino Acid Analysis

Na425

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na425

Lot Number: 4010001

Production Date: 02 Jan 2014

Expiration Date: 02 Jan 2016

Acceptance Criteria:

- | | |
|--|----------|
| 1. pH = 4.25 ± 0.01 @ 25 ± 0.4 °C (Y/N) | <u>y</u> |
| 2. Resolution meets specifications: | |
| • Alanine/Cystine (Y/N) | <u>y</u> |
| • Isoleucine/Leucine (Y/N) | <u>y</u> |
| 3. Elution time of Valine and Methionine meets specification (Y/N) | <u>y</u> |
| 4. Baseline: no peaks (Y/N) | <u>y</u> |

Calibration Standard: Protein Hydrolysate Standard, **PN:** 012506H

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** 24 Jan 2014
President
Title

Rev. A, 10/24/12

P:\Documentation\Chemical Documentation\Na425