



**CERTIFICATE OF ANALYSIS**  
**Sodium Eluant pH 4.25 for Amino Acid Analysis**

**Na425**

**Product:** Sodium Eluant for Amino Acid Analysis

**Part Number:** Na425

**Lot Number:** 405021

**Production Date:** 21 May 2014

**Expiration Date:** 21 May 2016

**Acceptance Criteria:**

1. pH =  $4.25 \pm 0.01$  @  $25 \pm 0.4$  °C (Y/N)

y

2. Resolution meets specifications:

• Alanine/Cystine (Y/N)

y

• Isoleucine/Leucine (Y/N)

y

3. Elution time of Valine and Methionine meets specification (Y/N)

y

4. Baseline: no peaks (Y/N)

y

**Calibration Standard:** Protein Hydrolysate Standard, **PN:** 012506H

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:**

Jim Murphy

**Date:**

09 Jun 2014

President

**Title**

Rev. A, 10/24/12

P:\Documentation\Chemical Documentation\Na425