



CERTIFICATE OF ANALYSIS
Sodium Eluant pH640 for Amino Acid Analysis
Na640

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na640

Lot Number: 208026

Production Date: 30 Aug 2012

Expiration Date: 30 Aug 2013

Acceptance Criteria:

1. pH = 6.40 ± 0.01 @ 25 ± 0.4 °C (Y/N) y
2. Resolution meets specifications: y
 - Lysine, Tryptophan, and Ammonia (Y/N)
3. Elution time of Histidine and Arginine meets specification (Y/N) y
4. Baseline: no peaks (Y/N) y

Calibration Standard: Protein Hydrolysate Standard, **PN:** 012506H

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** 12 Sep 2012
VP Operations
Title

Rev.8, 12/12/2011