



CERTIFICATE OF ANALYSIS
Sodium Eluant for Amino Acid Analysis
Na640

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na640

Lot Number: 004005

Test Date: 06 APR 2010


Acceptance Criteria:

1. pH = 6.40 ± 0.01 @ 25 ± 0.4 °C
2. Peak resolution meets specification

Actual Values

6.403

Y

Analysis Certified by: 
VP Ops
Title

Date: 13 Apr 2010

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 6 7/15/05