



**CERTIFICATE OF ANALYSIS**  
**Sodium Eluant for Amino Acid Analysis**  
**Na640**

**Product:** Sodium Eluant for Amino Acid Analysis

**Part Number:** Na640

**Lot Number:** 006025

**Test Date:** 18 JUN 2010

**Acceptance Criteria:**

1. pH =  $6.40 \pm 0.01$  @  $25 \pm 0.4$  °C
2. Peak resolution meets specification

**Actual Values**

6.397

Y

**Analysis Certified by:** 

**Date:** 23 JUN 2010

VP  
Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE