



CERTIFICATE OF ANALYSIS
Sodium Eluant for Amino Acid Analysis
Na640

Product: Sodium Eluant for Amino Acid Analysis
Part Number: Na640
Lot Number: 10602.5
Test Date: 24 Jun 2011


Acceptance Criteria:

1. pH = 6.40 ± 0.01 @ 25 ± 0.4 °C
2. Peak resolution meets specification

Actual Values

6.405
4

Analysis Certified by:


VP OPR
Title

Date:

28 Jun 2011

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE