



**CERTIFICATE OF ANALYSIS**  
**Sodium Eluant for Amino Acid Analysis**

**Na640**

**Product:** Sodium Eluant for Amino Acid Analysis  
**Part Number:** Na640  
**Lot Number:** 112028  
**Production Date:** 28 Dec 2011  
**Expiration Date:** 20 Dec 2012

**Acceptance Criteria:**

1. pH =  $6.40 \pm 0.01$  @  $25 \pm 0.4$  °C (Y/N)
2. Peak resolution meets specification (Y/N)

Y  
Y

**Sample:** Protein hydrolysate, PN: 012506H

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:**

*[Signature]*  
VP OPR  
Title

**Date:** 05 MAR 2012

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE