



CERTIFICATE OF ANALYSIS

Sodium Eluant for Amino Acid Analysis

Na740

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na740

Lot Number: 601010

Production Date: 12 Jan 2016

Expiration Date: 12 Jan 2018

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

- | | |
|---|----------|
| 1. pH > 7.40 @ 25 ± 0.4 °C (Y/N) | <u>y</u> |
| 2. Resolution meets specifications: | |
| • Tyrosine/Phenylalanine (Y/N) | <u>y</u> |
| 3. Retention time of L-Histidine meets specifications (Y/N) | <u>y</u> |
| 4. Baseline: no peaks (Y/N) | <u>y</u> |

Calibration Standard: Collagen Hydrolysate, **PN** 012506C,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** January 21, 2016
President
Title