



# CERTIFICATE OF ANALYSIS

## Sodium Eluant for Amino Acid Analysis

Na740

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na740

Lot Number: 003029

Test Start Date: 26 MAR 10

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH =  $7.55 \pm 0.06$  @  $25 \pm 0.4$  °C

Actual Values

7.585

2. Retention time of L-histidine meets specifications (Y/N).

Y

3. Resolution of Tyr/Phe  $\geq 75$  % (Y/N).

Y

Calibration Standard: Collagen Hydrolysate, PN 012506C, LN 007003

Analysis Certified by: \_\_\_\_\_

*Jim Murphy*  
VP OPR

Title

Date: \_\_\_\_\_

01 APR 10

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE