



# CERTIFICATE OF ANALYSIS

## Sodium Eluant for Amino Acid Analysis

Na740

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na740

Lot Number: 005013

Test Start Date: 18 MAY 2010

Primary Application: Sodium-based amino acid analysis.

**Acceptance Criteria:**

1. pH =  $7.55 \pm 0.06$  @  $25 \pm 0.4$  °C

**Actual Values**

7.520

2. Retention time of L-histidine meets specifications (Y/N).

Y

3. Resolution of Tyr/Phe  $\geq 75$  % (Y/N).

Y

Calibration Standard: Collagen Hydrolysate, PN 012506C, LN 002006

Analysis Certified by: \_\_\_\_\_

*Jim Young*

Date: 21 MAY 2010

*UP Operator*

Title

**ACCEPTED**

PICKERING LABORATORIES, INC.

QUALITY ASSURANCE